

Nederlands Meetinstituut

Certificate

Applicant : Instromet B.V.
Munstermanstraat 6
Silvolde, the Netherlands

Number : G1/S/4096rev1
Ordernr. : 312465
Vk nr. : 234365
Page : 1 of 2

Request for a : An electronic volume conversion device, Instromet, consisting of:
calibration of an EVCD, type 333, nr. 03545001 - 2004,
a temperature sensor with an operation range -10 °C to +40 °C,
an absolute pressure transmitter with a span of 1,5 bara
and an operation range from $p_{\min} = 0,9$ bara to $p_{\max} = 1,5$ bara.
From the volume V_1 at line conditions the EVCD calculates a volume V_n at the base
temperature $t_n = 0$ °C and at the base pressure $p_n = 1,01325$ bar, using


$$V_n = V_1 * p / p_n * (273,15 + t_n) / (273,15 + t) * C$$

where p is the absolute line pressure in bara and t is the line temperature in °C.
The value of the parameter $C = 1,0000$ is programmed as a preset value.
The EVCD is set to be used in combination with a gas meter that generates 1 impulse
per cubic meter indicated by the meter.

Test method : Volume impulses have been generated with an electrically gasmetersimulator.
The pressure readout of the EVCD has been compared with a deadweight tester.
The temperature readout of the EVCD has been compared with a digital reference
thermometer.
At $p = 1,2$ bara and $t = 15$ °C the operation of the system has been tested by counting
impulses from the gasmetersimulator. The indication V_1 of the EVCD has been
compared with the volume V_m indicated by the gasmetersimulator.

Silvolde, : August 17, 2004
NMI Certin B.V.

B. Pastoor
Afdeling Keuringen




NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
Telefoon (078) 6332 332
Telefax (078) 6332 309
E - Mail NMI@NMI.nl

Nederlands Meetinstituut N.V. (Registered at the
Chamber of Commerce Haaglanden number (27228701))

Subsidiary companies :
NMI Certin B.V. (27233418)
NMI Van Swinden Laboratorium B.V. (27233418)
NMI International B.V. (KvK nr. 27239176)

This certificate is issued under the provision that Nederlands
Meetinstituut N.V. nor its subsidiary companies accept any
liability.

Reproduction of the complete certificate is allowed.
Parts of the certificate may only be reproduced after written
Permission



Test method : The indication of V_n has been compared with the volume calculated from V_1 , the indicated values of p and t and C .

All testing equipment is traceable to international standards.
The ambient temperature during all tests was ± 20 °C.

Test date : May 5, 2004

Results : 1. Pressure transmitter.

Ref. press (bara)	Ind. press (bara)	Error %	Uncertainty %
1,4549	1,4550	0,01	0,10
1,2849	1,2850	0,01	0,10
0,9849	0,9850	0,01	0,15
1,1349	1,1350	0,01	0,10

2. Temperature sensor

Ref. temp. (K)	Ind. temp. (K)	Error (%)
263,25	263,24	0,00
313,07	313,03	-0,01
288,20	288,18	-0,01

The uncertainty in the error is about 0,05 %

3. Operation test at $p = 1,1350$ bara and $t = 15,03$ °C

a. Impulse transmission $V_m = V_1 = 150 * 1 \text{ m}^3$

b. Calculation

Indicated V_n	=	$159,2293 * 1 \text{ m}^3$
Calculated V_n	=	$159,2604 * 1 \text{ m}^3$
Resulting error	=	-0,02 %

