

CDM Instruments Maintenance Log

Sep. 2007 –Dec.2007

Date	Status																					
2007.09.10	<p>On-site check, no abnormal situation. Regular check of below instruments was conducted by site operator Mr. Ren Chuanhui. The result is as follows:</p> <table border="1"> <thead> <tr> <th>Before calibration</th> <th>Standards Gas</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>Inlet N2O</td> <td>Zero check Gas (N2)</td> <td>3 ppm</td> </tr> <tr> <td>Outlet N2O</td> <td>Zero check Gas (N2)</td> <td>0.9 ppm</td> </tr> <tr> <td>CH4</td> <td>Zero check Gas (N2)</td> <td>0.6 ppm</td> </tr> <tr> <td>Inlet N2O</td> <td>Span Check Gas 2760 ppm</td> <td>2762 ppm</td> </tr> <tr> <td>Outlet N2O</td> <td>Span Check Gas 432 ppm</td> <td>432 ppm</td> </tr> <tr> <td>CH4</td> <td>Span Check Gas 178 ppm</td> <td>178 ppm</td> </tr> </tbody> </table> <p>OK.</p>	Before calibration	Standards Gas	Value	Inlet N2O	Zero check Gas (N2)	3 ppm	Outlet N2O	Zero check Gas (N2)	0.9 ppm	CH4	Zero check Gas (N2)	0.6 ppm	Inlet N2O	Span Check Gas 2760 ppm	2762 ppm	Outlet N2O	Span Check Gas 432 ppm	432 ppm	CH4	Span Check Gas 178 ppm	178 ppm
Before calibration	Standards Gas	Value																				
Inlet N2O	Zero check Gas (N2)	3 ppm																				
Outlet N2O	Zero check Gas (N2)	0.9 ppm																				
CH4	Zero check Gas (N2)	0.6 ppm																				
Inlet N2O	Span Check Gas 2760 ppm	2762 ppm																				
Outlet N2O	Span Check Gas 432 ppm	432 ppm																				
CH4	Span Check Gas 178 ppm	178 ppm																				
2007.09.11	<p>On-site check, no abnormal situation. 10:40 start RCS; CDM facility was put into operation. Regular check of below instruments was conducted by site operator Mr. Ren Chuanhui. The result is as follows:</p> <table border="1"> <thead> <tr> <th>Before calibration</th> <th>Standards Gas</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>Inlet N2O</td> <td>Zero check Gas (N2)</td> <td>2 ppm</td> </tr> <tr> <td>Outlet N2O</td> <td>Zero check Gas (N2)</td> <td>0.4 ppm</td> </tr> <tr> <td>CH4</td> <td>Zero check Gas (N2)</td> <td>0.2 ppm</td> </tr> <tr> <td>Inlet N2O</td> <td>Span Check Gas 2760 ppm</td> <td>2761 ppm</td> </tr> <tr> <td>Outlet N2O</td> <td>Span Check Gas 432 ppm</td> <td>431 ppm</td> </tr> <tr> <td>CH4</td> <td>Span Check Gas 178 ppm</td> <td>179 ppm</td> </tr> </tbody> </table> <p>OK.</p>	Before calibration	Standards Gas	Value	Inlet N2O	Zero check Gas (N2)	2 ppm	Outlet N2O	Zero check Gas (N2)	0.4 ppm	CH4	Zero check Gas (N2)	0.2 ppm	Inlet N2O	Span Check Gas 2760 ppm	2761 ppm	Outlet N2O	Span Check Gas 432 ppm	431 ppm	CH4	Span Check Gas 178 ppm	179 ppm
Before calibration	Standards Gas	Value																				
Inlet N2O	Zero check Gas (N2)	2 ppm																				
Outlet N2O	Zero check Gas (N2)	0.4 ppm																				
CH4	Zero check Gas (N2)	0.2 ppm																				
Inlet N2O	Span Check Gas 2760 ppm	2761 ppm																				
Outlet N2O	Span Check Gas 432 ppm	431 ppm																				
CH4	Span Check Gas 178 ppm	179 ppm																				
2007.09.16	<p>Regular on-site check, all the instruments in good condition. Mr. Yang Junwei from ABB conducted the regular check of below instruments. The result is as follows:</p> <table border="1"> <thead> <tr> <th>Before calibration</th> <th>Standards Gas</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>INLET N2O</td> <td>Zero Check Gas (N2)</td> <td>5 ppm</td> </tr> <tr> <td>OUTLET N2O</td> <td>Zero Check Gas (N2)</td> <td>1.2 ppm</td> </tr> <tr> <td>CH4</td> <td>Zero Check Gas (N2)</td> <td>-0.3 ppm</td> </tr> <tr> <td>INLET N2O</td> <td>Span Check Gas 2760 ppm</td> <td>2763 ppm</td> </tr> <tr> <td>OUTLET N2O</td> <td>Span Check Gas 432 ppm</td> <td>431 ppm</td> </tr> <tr> <td>CH4</td> <td>Span Check Gas 178 ppm</td> <td>178 ppm</td> </tr> </tbody> </table> <p>OK.</p>	Before calibration	Standards Gas	Value	INLET N2O	Zero Check Gas (N2)	5 ppm	OUTLET N2O	Zero Check Gas (N2)	1.2 ppm	CH4	Zero Check Gas (N2)	-0.3 ppm	INLET N2O	Span Check Gas 2760 ppm	2763 ppm	OUTLET N2O	Span Check Gas 432 ppm	431 ppm	CH4	Span Check Gas 178 ppm	178 ppm
Before calibration	Standards Gas	Value																				
INLET N2O	Zero Check Gas (N2)	5 ppm																				
OUTLET N2O	Zero Check Gas (N2)	1.2 ppm																				
CH4	Zero Check Gas (N2)	-0.3 ppm																				
INLET N2O	Span Check Gas 2760 ppm	2763 ppm																				
OUTLET N2O	Span Check Gas 432 ppm	431 ppm																				
CH4	Span Check Gas 178 ppm	178 ppm																				
2007.09.22	<p>Regular on-site check, all the instruments in good condition. Mr. Yang Junwei from ABB conducted the regular check of below instruments. The result is as follows:</p>																					

		<table border="1"> <thead> <tr> <th>Before calibration</th> <th>Standards Gas</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>INLET N2O</td> <td>Zero Check Gas (N2)</td> <td>2 ppm</td> </tr> <tr> <td>OUTLET N2O</td> <td>Zero Check Gas (N2)</td> <td>0.4 ppm</td> </tr> <tr> <td>CH4</td> <td>Zero Check Gas (N2)</td> <td>0.2 ppm</td> </tr> <tr> <td>INLET N2O</td> <td>Span Check Gas 2760 ppm</td> <td>2761 ppm</td> </tr> <tr> <td>OUTLET N2O</td> <td>Span Check Gas 432 ppm</td> <td>431 ppm</td> </tr> <tr> <td>CH4</td> <td>Span Check Gas 178 ppm</td> <td>179 ppm</td> </tr> </tbody> </table>	Before calibration	Standards Gas	Value	INLET N2O	Zero Check Gas (N2)	2 ppm	OUTLET N2O	Zero Check Gas (N2)	0.4 ppm	CH4	Zero Check Gas (N2)	0.2 ppm	INLET N2O	Span Check Gas 2760 ppm	2761 ppm	OUTLET N2O	Span Check Gas 432 ppm	431 ppm	CH4	Span Check Gas 178 ppm	179 ppm
Before calibration	Standards Gas	Value																					
INLET N2O	Zero Check Gas (N2)	2 ppm																					
OUTLET N2O	Zero Check Gas (N2)	0.4 ppm																					
CH4	Zero Check Gas (N2)	0.2 ppm																					
INLET N2O	Span Check Gas 2760 ppm	2761 ppm																					
OUTLET N2O	Span Check Gas 432 ppm	431 ppm																					
CH4	Span Check Gas 178 ppm	179 ppm																					
	OK.																						
2007.09.28	<p>Regular on-site check, all the instruments in good condition. Site operator Mr. Ren Chuanhui conducted the regular check of below instruments. The result is as follows:</p>	<table border="1"> <thead> <tr> <th>Before calibration</th> <th>Standards Gas</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>INLET N2O</td> <td>Zero Check Gas (N2)</td> <td>4 ppm</td> </tr> <tr> <td>OUTLET N2O</td> <td>Zero Check Gas (N2)</td> <td>0.1 ppm</td> </tr> <tr> <td>CH4</td> <td>Zero Check Gas (N2)</td> <td>0.5 ppm</td> </tr> <tr> <td>INLET N2O</td> <td>Span Check Gas 2760 ppm</td> <td>2762 ppm</td> </tr> <tr> <td>OUTLET N2O</td> <td>Span Check Gas 432 ppm</td> <td>433 ppm</td> </tr> <tr> <td>CH4</td> <td>Span Check Gas 178 ppm</td> <td>177 ppm</td> </tr> </tbody> </table>	Before calibration	Standards Gas	Value	INLET N2O	Zero Check Gas (N2)	4 ppm	OUTLET N2O	Zero Check Gas (N2)	0.1 ppm	CH4	Zero Check Gas (N2)	0.5 ppm	INLET N2O	Span Check Gas 2760 ppm	2762 ppm	OUTLET N2O	Span Check Gas 432 ppm	433 ppm	CH4	Span Check Gas 178 ppm	177 ppm
Before calibration	Standards Gas	Value																					
INLET N2O	Zero Check Gas (N2)	4 ppm																					
OUTLET N2O	Zero Check Gas (N2)	0.1 ppm																					
CH4	Zero Check Gas (N2)	0.5 ppm																					
INLET N2O	Span Check Gas 2760 ppm	2762 ppm																					
OUTLET N2O	Span Check Gas 432 ppm	433 ppm																					
CH4	Span Check Gas 178 ppm	177 ppm																					
	OK.																						
2007.10.08	<p>Regular on-site check, DCS in good condition; RCS Malfunction. Mr. Yang Junwei from ABB conducted the regular check of below instruments. The result is as follows:</p>	<table border="1"> <thead> <tr> <th>Before calibration</th> <th>Standards Gas</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>INLET N2O</td> <td>Zero Check Gas (N2)</td> <td>1 ppm</td> </tr> <tr> <td>OUTLET N2O</td> <td>Zero Check Gas (N2)</td> <td>0.6 ppm</td> </tr> <tr> <td>CH4</td> <td>Zero Check Gas (N2)</td> <td>0.2 ppm</td> </tr> <tr> <td>INLET N2O</td> <td>Span Check Gas 2760 ppm</td> <td>2760 ppm</td> </tr> <tr> <td>OUTLET N2O</td> <td>Span Check Gas 432 ppm</td> <td>433.4 ppm</td> </tr> <tr> <td>CH4</td> <td>Span Check Gas 178 ppm</td> <td>178 ppm</td> </tr> </tbody> </table>	Before calibration	Standards Gas	Value	INLET N2O	Zero Check Gas (N2)	1 ppm	OUTLET N2O	Zero Check Gas (N2)	0.6 ppm	CH4	Zero Check Gas (N2)	0.2 ppm	INLET N2O	Span Check Gas 2760 ppm	2760 ppm	OUTLET N2O	Span Check Gas 432 ppm	433.4 ppm	CH4	Span Check Gas 178 ppm	178 ppm
Before calibration	Standards Gas	Value																					
INLET N2O	Zero Check Gas (N2)	1 ppm																					
OUTLET N2O	Zero Check Gas (N2)	0.6 ppm																					
CH4	Zero Check Gas (N2)	0.2 ppm																					
INLET N2O	Span Check Gas 2760 ppm	2760 ppm																					
OUTLET N2O	Span Check Gas 432 ppm	433.4 ppm																					
CH4	Span Check Gas 178 ppm	178 ppm																					
	OK.																						
14/10/2007 - 23/11/2007	- CDM facility is under downtime.																						
2007.11.20	<p>From 14/10/2007 to 23/11/2007, CDM facility is under downtime. Mr. Yang Junwei from ABB conducted the regular check of below instruments. The result is as follows:</p>	<table border="1"> <thead> <tr> <th>Before calibration</th> <th>Standards Gas</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>INLET N2O</td> <td>Zero Check Gas (N2)</td> <td>6 ppm</td> </tr> <tr> <td>OUTLET N2O</td> <td>Zero Check Gas (N2)</td> <td>0.8 ppm</td> </tr> <tr> <td>CH4</td> <td>Zero Check Gas (N2)</td> <td>0.1 ppm</td> </tr> <tr> <td>INLET N2O</td> <td>Span Check Gas 2760 ppm</td> <td>2764 ppm</td> </tr> <tr> <td>OUTLET N2O</td> <td>Span Check Gas 432 ppm</td> <td>431 ppm</td> </tr> <tr> <td>CH4</td> <td>Span Check Gas 178 ppm</td> <td>179 ppm</td> </tr> </tbody> </table>	Before calibration	Standards Gas	Value	INLET N2O	Zero Check Gas (N2)	6 ppm	OUTLET N2O	Zero Check Gas (N2)	0.8 ppm	CH4	Zero Check Gas (N2)	0.1 ppm	INLET N2O	Span Check Gas 2760 ppm	2764 ppm	OUTLET N2O	Span Check Gas 432 ppm	431 ppm	CH4	Span Check Gas 178 ppm	179 ppm
Before calibration	Standards Gas	Value																					
INLET N2O	Zero Check Gas (N2)	6 ppm																					
OUTLET N2O	Zero Check Gas (N2)	0.8 ppm																					
CH4	Zero Check Gas (N2)	0.1 ppm																					
INLET N2O	Span Check Gas 2760 ppm	2764 ppm																					
OUTLET N2O	Span Check Gas 432 ppm	431 ppm																					
CH4	Span Check Gas 178 ppm	179 ppm																					

	OK.																														
2007.12.07	<p>From 24/11/2007 to 08/12/2007, CDM facility is under downtime. Mr. Yang Junwei from ABB conducted the regular check of below instruments. The result is as follows:</p> <table border="1"> <thead> <tr> <th>Before calibration</th> <th>Standards Gas</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>INLET N2O</td> <td>Zero Check Gas (N2)</td> <td>2 ppm</td> </tr> <tr> <td>OUTLET N2O</td> <td>Zero Check Gas (N2)</td> <td>0.7 ppm</td> </tr> <tr> <td>CH4</td> <td>Zero Check Gas (N2)</td> <td>0.3 ppm</td> </tr> <tr> <td>INLET N2O</td> <td>Span Check Gas 2760 ppm</td> <td>2762 ppm</td> </tr> <tr> <td>OUTLET N2O</td> <td>Span Check Gas 432 ppm</td> <td>434 ppm</td> </tr> <tr> <td>CH4</td> <td>Span Check Gas 178 ppm</td> <td>178 ppm</td> </tr> </tbody> </table> <p>OK.</p>	Before calibration	Standards Gas	Value	INLET N2O	Zero Check Gas (N2)	2 ppm	OUTLET N2O	Zero Check Gas (N2)	0.7 ppm	CH4	Zero Check Gas (N2)	0.3 ppm	INLET N2O	Span Check Gas 2760 ppm	2762 ppm	OUTLET N2O	Span Check Gas 432 ppm	434 ppm	CH4	Span Check Gas 178 ppm	178 ppm									
Before calibration	Standards Gas	Value																													
INLET N2O	Zero Check Gas (N2)	2 ppm																													
OUTLET N2O	Zero Check Gas (N2)	0.7 ppm																													
CH4	Zero Check Gas (N2)	0.3 ppm																													
INLET N2O	Span Check Gas 2760 ppm	2762 ppm																													
OUTLET N2O	Span Check Gas 432 ppm	434 ppm																													
CH4	Span Check Gas 178 ppm	178 ppm																													
2007.12.15	<p>Mr. Yang Junwei from ABB conducted the regular check of below instruments. The result is as follows:</p> <table border="1"> <thead> <tr> <th>Before calibration</th> <th>Standards Gas</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>INLET N2O</td> <td>Zero Check Gas (N2)</td> <td>10 ppm</td> </tr> <tr> <td>OUTLET N2O</td> <td>Zero Check Gas (N2)</td> <td>0.2 ppm</td> </tr> <tr> <td>CH4</td> <td>Zero Check Gas (N2)</td> <td>0.6 ppm</td> </tr> <tr> <td>INLET N2O</td> <td>Span Check 2760 ppm</td> <td>2769 ppm</td> </tr> <tr> <td>OUTLET N2O</td> <td>Span Check 432 ppm</td> <td>452 ppm</td> </tr> <tr> <td>CH4</td> <td>Span Check 178 ppm</td> <td>444 ppm</td> </tr> </tbody> </table> <p>OUTLET N2O and CH4 are OK. Check INLET N2O, the result is as follows:</p> <table border="1"> <thead> <tr> <th>Before calibration</th> <th>Standards Gas</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>INLET N2O</td> <td>Zero Check Gas (N2)</td> <td>0 ppm</td> </tr> <tr> <td>INLET N2O</td> <td>Span Check Gas 2760 ppm</td> <td>2760 ppm</td> </tr> </tbody> </table>	Before calibration	Standards Gas	Value	INLET N2O	Zero Check Gas (N2)	10 ppm	OUTLET N2O	Zero Check Gas (N2)	0.2 ppm	CH4	Zero Check Gas (N2)	0.6 ppm	INLET N2O	Span Check 2760 ppm	2769 ppm	OUTLET N2O	Span Check 432 ppm	452 ppm	CH4	Span Check 178 ppm	444 ppm	Before calibration	Standards Gas	Value	INLET N2O	Zero Check Gas (N2)	0 ppm	INLET N2O	Span Check Gas 2760 ppm	2760 ppm
Before calibration	Standards Gas	Value																													
INLET N2O	Zero Check Gas (N2)	10 ppm																													
OUTLET N2O	Zero Check Gas (N2)	0.2 ppm																													
CH4	Zero Check Gas (N2)	0.6 ppm																													
INLET N2O	Span Check 2760 ppm	2769 ppm																													
OUTLET N2O	Span Check 432 ppm	452 ppm																													
CH4	Span Check 178 ppm	444 ppm																													
Before calibration	Standards Gas	Value																													
INLET N2O	Zero Check Gas (N2)	0 ppm																													
INLET N2O	Span Check Gas 2760 ppm	2760 ppm																													
2007.12.17-19	TÜV SÜD do QAL2 (including zero and span checks etc.)																														
2007.12.28	<p>15:00, no power, from 18:00, CDM facility is put into action again, regular check of below instruments. The result is as follows:</p> <table border="1"> <thead> <tr> <th>Before calibration</th> <th>Standards Gas</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>INLET N2O</td> <td>Zero Check Gas (N2)</td> <td>14 ppm</td> </tr> <tr> <td>OUTLET N2O</td> <td>Zero Check Gas (N2)</td> <td>-0.1 ppm</td> </tr> <tr> <td>CH4</td> <td>Zero Check Gas (N2)</td> <td>0.3 ppm</td> </tr> <tr> <td>INLET N2O</td> <td>Span Check 2760 ppm</td> <td>2773 ppm</td> </tr> <tr> <td>OUTLET N2O</td> <td>Span Check 432 ppm</td> <td>452 ppm</td> </tr> <tr> <td>CH4</td> <td>Span Check 178 ppm</td> <td>445 ppm</td> </tr> </tbody> </table> <p>OUTLET N2O and CH4 are OK. Check INLET N2O, the result is as follows:</p> <table border="1"> <thead> <tr> <th>Before calibration</th> <th>Standards Gas</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>INLET N2O</td> <td>Zero Check Gas (N2)</td> <td>0 ppm</td> </tr> <tr> <td>INLET N2O</td> <td>Span Check Gas 2760 ppm</td> <td>2760 ppm</td> </tr> </tbody> </table>	Before calibration	Standards Gas	Value	INLET N2O	Zero Check Gas (N2)	14 ppm	OUTLET N2O	Zero Check Gas (N2)	-0.1 ppm	CH4	Zero Check Gas (N2)	0.3 ppm	INLET N2O	Span Check 2760 ppm	2773 ppm	OUTLET N2O	Span Check 432 ppm	452 ppm	CH4	Span Check 178 ppm	445 ppm	Before calibration	Standards Gas	Value	INLET N2O	Zero Check Gas (N2)	0 ppm	INLET N2O	Span Check Gas 2760 ppm	2760 ppm
Before calibration	Standards Gas	Value																													
INLET N2O	Zero Check Gas (N2)	14 ppm																													
OUTLET N2O	Zero Check Gas (N2)	-0.1 ppm																													
CH4	Zero Check Gas (N2)	0.3 ppm																													
INLET N2O	Span Check 2760 ppm	2773 ppm																													
OUTLET N2O	Span Check 432 ppm	452 ppm																													
CH4	Span Check 178 ppm	445 ppm																													
Before calibration	Standards Gas	Value																													
INLET N2O	Zero Check Gas (N2)	0 ppm																													
INLET N2O	Span Check Gas 2760 ppm	2760 ppm																													

负责人 (Functionary):

(Chuanhui Ren 任传辉)