



GUIDANCE RELATED TO CALIBRATION (MONITORING) REQUIREMENTS

(Extract of the report of the twenty-fourth meeting of the Executive Board, paragraph 37)

The Board accepted the Meth Panel's recommendation that a zero check cannot be considered as a substitute for calibration of the measurement instrument.